

File: No

Integration Events: AutoIntegrate

Generate Report During Run Method: No

Signal Correlation Window: 0.020

Quantitative Report Settings

Report Type: Summary

Output Destination

Screen: Yes

Printer: No

File: No

Generate Report During Run Method: No

Quantification Testostérone / Epitestostérone
Calibration Last Updated: Fri Jul 21 16:01:39 2006

Reference Window: 2.00 Minutes
Non-Reference Window: 1.00 Minutes
Correlation Window: 0.10 minutes
Default Multiplier: 1.00
Default Sample Concentration: 0.00

Compound Information

-
1) Méthyltestostérone (ISTD TR)
Ret. Time 20.91 min., Extract & Integrate from 20.41 to 21.41 min.

Lvl ID	Conc (ng/ml)	Response
1	100.000	3723520
2	100.000	3989786
3	100.000	5093742

ISTD conc: 100.000 ng/ml
Curve Fit: Linear

-
2) Epitestostérone ()
Ret. Time 18.50 min., Extract & Integrate from 18.00 to 19.00 min.

Lvl ID	Conc (ng/ml)	Response
1	5.000	149154
2	30.000	1136908
3	60.000	3173258

Curve Fit: Linear, forced through origin
